



CERTIFICATE OF ANALYSIS

PROPANE

INSTRUMENT PURITY

Guaranteed Specifications

99.5 % +

Propane

Sum of all impurities to

Methane

be < 0.50 %

Ethane

Water < 5 PPM

Propylene

Isobutane

Liquid Phase Analysis

n-Butane

NeoPentane

IsoPentane

n-Pentane

C6+

Hexane

BTEX

Carbon Dioxide

Water

Oxygen

Nitrogen

Sulfur

Lot Analysis

99.70 % +

< 1 PPM

7 PPM

12 PPM

2526 PPM

385 PPM

44 PPM

61 PPM

< 1 PPM

Non Detected

Non Detected

Non Detected

< 1 PPM

< 5 PPM

< 5 PPM

< 20 PPM

Non Detected

“This document certifies that the contents of this finished goods container conform to the specifications listed.

This confirmation is based upon analysis of the source raw material and does not reflect analysis of the finished goods container.”

Batch Number
PRO-103124-I

Please consult the SDS for further information.

BVV represents only that the Product shall meet the specifications herein. All transactions involving this product are subject to BVV standard terms and conditions. BVV makes no additional representations or warranties, express or implied, through this COA.