



1251 Frontenac Rd., Naperville, IL 60563
Phone: 331-281-0154 e-mail: support@shopbvv.com
www.shopbvv.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: ULTRAPURE RODI WATER USP - PAGE 1
LOT: 01111BSMF
MANUFACTURE DATE: 02/11/2026

TEST PROCEDURE:	METHOD:	SPECIFICATIONS:	RESULTS:
RESISTIVITY MEGOHM-CM	10,040.060	> 16.5	18.0
TOC PPB	14,000.020	< 500	< 500
CONDUCTIVITY MICROMHO	10,040.060	VESSEL 1 < 1.0	0.40
		VESSEL 2 < 1.0	0.52
		VESSEL 3 < 1.0	0.49
		VESSEL 4 < 1.0	0.69
		VESSEL 5 < 1.0	0.51
		VESSEL 6 < 1.0	0.56
		VESSEL 7 < 1.0	0.47

Please consult the SDS for further information.

BVV Guarantees only that the product describe herein meets the specifications outlined in this document. All Transactions involving this product are governed by BVV's standard terms and conditions. This Certificate of Analysis (COA) does not include any additional information, guarantees, or warranties, whether express or implied beyond those explicitly stated.



1251 Frontenac Rd., Naperville, IL 60563
Phone: 331-281-0154 e-mail: support@shopbvv.com
www.shopbvv.com

CERTIFICATE OF ANALYSIS

PRODUCT NAME: ULTRAPURE RODI WATER USP - PAGE 2
LOT: 01111BSMF
MANUFACTURE DATE: 02/11/2026

TEST PROCEDURE:	METHOD:	SPECIFICATIONS:	RESULTS:
CONDUCTIVITY MICROMHO	10.040.060	VESSEL 8 < 1.0	0.44
		VESSEL 9 < 1.0	0.47
		VESSEL 10 < 1.0	0.48
		VESSEL 11 < 1.0	0.34
		VESSEL 12 < 1.0	0.48
		VESSEL 13 < 1.0	0.42
		VESSEL 14 < 1.0	0.37
		VESSEL 15 < 1.0	0.52
		VESSEL 16 < 1.0	0.31

Please consult the SDS for further information.

BVV Guarantees only that the product describe herein meets the specifications outlined in this document. All Transactions involving this product are governed by BVV's standard terms and conditions. This Certificate of Analysis (COA) does not include any additional information, guarantees, or warranties, whether express or implied beyond those explicitly stated.